Round Robin

Quality of testing?

You can expect it, if you inspect it!






EN ISO 17025 / EN ISO/IEC 17025 / ISO-Guide 43-1

EN ISO 17025:

Participation in programmes to compare results of testmethods with other laboratories



The "proficiency test" is a tool to determine the influence of parameters which may vary between individual laboratories, it does not represent a substitute for the calibration procedure, the use of calibrated equipment is taken for granted.





Basics

- All tests shall be performed under repeatability conditions (same operator, same equipment, in a short intervall of time).
- Statistical evaluations possess a weakened explanatory force when few labs participate in the evaluation.
- Extremely biased results have to be investigated for possible errors.
- Results with a z-score of more than 10 are discarded to prevent an unnecessary distortion of the evaluation (check with participant) and to prevent a too "optimistic" performance assessment of the other participants.
- Laboratory performance: expressed by laboratories bias (deviation of the labresult form assigned value X; X determined as a consensus value from participants, x*: "robust average"



Basics

- z-score: measure of the distance of an individual result from the mean, scale unit: standard deviation: gives a measure of the lab`s-performance.
- Identification of stragglers and outliers (Cochran`s and Grubb`s test); Grubb`s outlier: z-score > 2
- z-score > 2 denotes that the result of the respective lab deviates by more than \pm +/- 2σ from the accepted reference value for the proficiency assessment X. It is considered as a "warning signal". Approx. 95% of all results may lie in the intervall of $X \pm$ -/- 2σ if data is normally distributed.
- z-score of more than 3 shall be considered as an "action signal"
- Check for normal distribution of data by using the Anderson-Darling Test (after elimination of outliers)




Basics

- Reliable information on the method accuracy and lab performance depends on the limit of participants (min. 7). Other conditions to be met: standard uncertainty u_x max. 0,3 x s*, number of participants p* min. 16, data comes from a normal distribution
- Calculation of repeatability and reproducibility: for utilization of the measurement uncertainty
- Evaluated reproducibility limit R matters when two independent test results are compared; generally two results shall be judged not to be different if they differ by less than R (important for estimation of measurement uncertainty MU).



Basics

 Split level sampling for multiple check of test method accuracy as well as multiple check of the lab-performance; analysis of random vs systematic error: Youden plot (x-axis: level 1 response value / y-axis: level 2 response value).





ofi PTS 2008 – Test method: Abrasion resistance of rubber surface using Taber Abrader – ISO 5470-1





Taber Abrader – ISO 5470-1 - Evaluation

- Test procedure
- Results submitted and calculated values
- Graphical presentation of the test results (line plot)
- Evaluation of the laboratory performance (Bar chart)
- Additional check of the test method accuracy (table)
- Youden plot (Scatter diagram)



Test procedure

- Conditioning: 23°C/50%
- Taber Abrader: device acc. ISO 5470-1 (labs using devices acc. DIN 53754 had to adapt their device, device parameter had to be reported)

LabCodeNo.	3	26	33	37	112	167
Device compliance ^a)	ISO	DIN ⁵)	DIN	DIN	ISO ^b)	ISO
Distance between rotation axes	19,1 mm	19,1 mm	20 mm	not	20 mm	19,1 mm
(wheel / specimen)				reported		
ID of the nozzle	团	团	20 at 20		Ø	Ø
Nozzle distance from the specimen	Ø	Ø	Ø	~ ~ ~	Ø	Ø
Air pressure	Ø	M	107 124 94		Ø	")
Load	Ø	Ø	\square	V	Ø	Ø
Diameter of abrasive wheels	Ø	図	Ø	Ø	Ø	Ø
Width of the abrasive wheels	Ø	Ø	Ø	Ø	Ø	Ø

The testing device complies in its original state with ISO 5470-1 or with DIN 53754

- Test procedure acc. ISO 5470-1
- Expression of results acc. ISO 5470-1: mg/100 revolutions



This may be an erroneous indication because the standard distance between rotation axes (wheel / specimen) is 20 mm in devices which comply with the requirements according to DIN, and 19,1 mm in devices which comply with the requirements according to ISO

²⁾ LabCodeNo. 167, air pressure may comply but it was not checked

Results submitted & calculated values

Abrasion resistance of rubber surface Δm (using Taber abrader) - Sample A ISO 5470-1

₹esu	ılts sub	mitted b	y part	icipants		and		Results of	of robust	statis	tics	
i.e., individual results x_{ik} + number of the test repetitions made by each lab (n_{i})						(bottom part of the table)						
+ with	in laborat	e test repetiti ory means (, s for outliers	x_i) and			· _i)		Number of re Number of re	eporting labo	oratorie results	es p *: s $_{\Sigma}n_{i}$:	1:
Lab					Statistical evaluation of the			Outliers				
Code No.	Test replication No. (k) 1 2 3 4 5					6	submitted test results x_{ik} n_i x_i s_i			Cochran	Grubbs	1º
26 167 33 37 37 37 37 37 37 37 37 37 37 37 37	16,43 38,50 41,33 39,67 55,21 99,69	21.90 37.00 40.63 42.60 61.40 92.50					2 2 2 2 2 2	19,167 37,750 40,983 41,133 58,307 96,097	3,8655 1,0607 0,4950 2,0742 4,3746 5,0865			***************************************
					Robi	ıst average:	x* =	43.5	← assigne	d value	for t	16
	Robi	ust standar	d devia	ition for the		assessment			proficien			
Nur		epeate mea	asurem	ents neces	sary due t	o <i>s ,.ls</i> * -ratio	n' =	1	ОК	see p	age 4 f	or th
		Sta	ndard u	ıncertainty	of the ass	signed value:	$u_x =$	9,08745	NOT OK		ng of N	




Accuracy of test method


Do the input data come from a normal distribution (when outliers found were eliminated) ? (The results listed below shall be considered as really justified only if the input data come from a normal distribution)						
General mean $\sum n_i x_{ik} / \sum n_i$	m	39,5	mg			
Repeatability variance	S _F ²	7,9503733				
Repeatability standard deviation	S_r	2,81964	mg			
Repeatability coefficient of variation	CV%,	7,144	%			
Between-laboratory variance	S _L 2	189,7899611				
Between-laboratory standard deviation	s_L	13,77643	mg			
Between-laboratory coefficient of variation	$CV\%_L$	34,905	%			
Reproducibility variance s_R^2	$s_{r}^{2} + s_{L}^{2}$	197,7403344				
Reproducibility standard deviation	S _R	14,06202	mg			
Reproducibility coefficient of variation	$CV\%_R$	35,629	%			
Repeatability limit	r	7,9	mg			
Relative repeatability limit	r _{rel}	20	%			
Reproducibility limit	R	39,4	mg			
Relative reproducibility limit	R_{rel}	100	%			
Number of participants included in the accuracy evaluation	р	5	•			
Number of tests included in the accuracy evaluation	Σ^n	10				



Performance of individual laboratory


Abrasion resistance of rubber surface Δm (using Taber abrader) - Sample A







Youden plot confidence ellipse





ofi PTS 2008; January – June 2008

- 180 testing laboratories of 33 countries (from all continents)
- 80 test-methods were offered, mainly acc. to ISO- and EN-standard
- Material testing: plastic and rubber
- Product testing: geotextiles and geosynthetics, plastic pipes, surfaces and coatings, plastic films and packaging material, plastic sheeting, rigid foam and sound-insulating material, material for car interiors (testing for emissions)



• *ofi* PTS 2009

Testing methods for sports surfaces?



